

Declaration of Conformity

(according to Regulation (EU) 2017/745 on medical devices, annex IV dated May 2017)

CMD GmbH
AT-MF-000020244
Gewerbepark 5
4652 Fischlham
Austria

declares under its sole responsibility that the product


42604264814299WO-BN9
Skin-Elast-Hook

The purpose of the retractors is to spread the surgical wound and hold tissue away to give the surgeon better access to the surgical site.

Serial numbers 14299W0-B, 14299W0-G, 15097W0-B, 15097W0-G
valid until **07.04.2030**

to which this declaration relates is in conformity with the following regulation:
Evaluation procedure according to Regulation (EU) 2017/745 on medical devices, annex IX
Classification according to Regulation (EU) 2017/745 on medical devices, annex VIII:
medical device class IIa according to classification rule 7 (basic rule)

This product has been CE-labelled due to the fulfilment of the general safety and performance requirements according to Regulation (EU) 2017/745 on medical devices, annex I.



Ente Certificazione Macchine s.r.l.
Via Ca' Bella, 243/A – 40053 Valsamoggia (BO) – Italia

CMD GmbH
Fischlham, 11.04.2025

Christian **HEFNER**
Management