Declaration of Conformity

(according to Regulation (EU) 2017/745 on medical devices, annex IV dated May 2017)

CMD GmbH
AT-MF-000020244
Gewerbepark 5
4652 Fischlham
Austria

declares under its sole responsibility that the product

42604264814299WO-BN9
Skin-Elast-Hook

The purpose of the retractors is to spread the surgical wound and hold tissue away to give the surgeon better access to the surgical site.

Serial numbers 14299W0-B, 14299W0-G, 15097W0-B, 15097W0-G
valid until **07.04.2030**

to which this declaration relates is in conformity with the following regulation:

Evaluation procedure according to Regulation (EU) 2017/745 on medical devices, annex IX

Classification according to Regulation (EU) 2017/745 on medical devices, annex VIII:

medical device class IIa according to classification rule 7 (basic rule)

This product has been CE-labelled due to the fulfilment of the general safety and performance requirements according to Regulation (EU) 2017/745 on medical devices, annex I.

CE 2962

QMD Services GmbH
Zelinkagasse 10/3 – 1010 Vienna – Austria

CMD GmbH
Fischlham, 09.11.2026

Christian **HEFNER**
Management

Christian Hefner
Digital signiert von Christian Hefner
DN: C=AT, CN=Christian Hefner, O=CMD Medical GmbH, E=christian.hefner@cmd-medical.at
Datum: 2026.11.09 08:26:33 +01'00'

Printed version is not subject to change management. Always use the up-to-date document.